

ADDENDUM #1
RFP# 23-05

NMR Spectrometer

May 30, 2023

QUESTIONS AND ANSWERS

This Addendum contains one (1) page.

- 1) What nuclei are needed to be observed for a successful proposal?
a.) ^1H , ^{19}F & ^{13}C or comparable.
- 2) Is there a sensitivity requirement for a successful proposal?
a.) None specific but should have at least standard ^1H or other most common ones.
- 3) What is the minimum resolution required for a successful proposal?
a.) Better than 18ppb
- 4) Is the RFP committee interested in any customer references? If so, how can we provide contact information?
a.) Yes, please see the Business Reference Form in the Required Forms packet.
- 5) Does the end user have any specific nuclei requirements for this instrument?
a.) No
- 6) Would it be possible for Holyoke Community College to accept our submission electronically via email rather than a mailed hardcopy?
a.) Yes, the submissions may be submitted by email to bjackson@hcc.edu clearly marked “Bid on Requisition NO. 23-05 Enclosed”.

***** END OF ADDENDUM #1 *****